

# Accreditation



The Deutsche Akkreditierungsstelle attests with this **Accreditation Certificate** that the calibration laboratory

**JENOPTIK Industrial Metrology Germany GmbH**  
**Drachenloch 5, 78052 Villingen-Schwenningen**

meets the requirements according to DIN EN ISO/IEC 17025:2018 for the conformity assessment activities listed in the annex to this certificate. This includes additional existing legal and normative requirements for the calibration laboratory, including those in relevant sectoral schemes, provided they are explicitly confirmed in the annex to this certificate.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

This accreditation was issued in accordance with Art. 5 Para. 1 Sentence 2 of Regulation (EC) 765/2008, after an accreditation procedure was carried out in compliance with the minimum requirements of DIN EN ISO/IEC 17011 and on the basis of a review and decision of the appointed accreditation committees.

This accreditation certificate only applies in connection with the notices of 17.03.2025 with accreditation number D-K-15030-01.

It consists of this cover sheet, the reverse side of the cover sheet and the following annex with a total of 6 pages.

Registration number of the accreditation certificate: **D-K-15030-01-00**

Berlin, 17.03.2025

Dr. Florian Witt  
Head of Technical Unit

Translation issued:  
17.03.2025

  
Dr. Florian Witt  
Head of Technical Unit

*The certificate together with the annex reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH ([www.dakks.de](http://www.dakks.de)).*

This document is a translation. The definitive version is the original German accreditation certificate.

See notes overleaf

# Deutsche Akkreditierungsstelle GmbH

Office Berlin  
Spittelmarkt 10  
10117 Berlin

Office Frankfurt am Main  
Europa-Allee 52  
60327 Frankfurt am Main

Office Braunschweig  
Bundesallee 100  
38116 Braunschweig

The Deutsche Akkreditierungsstelle GmbH (DAkKS) is the entrusted national accreditation body of the Federal Republic of Germany according to § 8 section 1 AkkStelleG in conjunction with § 1 section 1 AkkStelleGBV. DAkKS is designated as the national accreditation authority by Germany according to Art. 4 Para. 4 of Regulation (EC) 765/2008 and clause 4.7 of DIN EN ISO/IEC 17000.

Pursuant to Art. 11 section 2 of Regulation (EC) 765/2008, the accreditation certificate shall be recognised as equivalent by the national authorities within the scope of this Regulation as well as by the WTO member states that have committed themselves in bilateral or multilateral mutual agreements to recognise the certificates of accreditation bodies that are members of ILAC or IAF as equivalent.

DAkKS is a signatory to the multilateral agreements for mutual recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Co-operation (ILAC).

The up-to-date state of membership can be retrieved from the following websites:

EA: [www.european-accreditation.org](http://www.european-accreditation.org)

ILAC: [www.ilac.org](http://www.ilac.org)

IAF: [www.iaf.nu](http://www.iaf.nu)

## Deutsche Akkreditierungsstelle

### Annex to the Accreditation Certificate D-K-15030-01-00 according to DIN EN ISO/IEC 17025:2018

**Valid from:** 17.03.2025

**Date of issue:** 17.03.2025

Holder of accreditation certificate:

**JENOPTIK Industrial Metrology Germany GmbH  
Drachenloch 5, 78052 Villingen-Schwenningen**

with the location

**JENOPTIK Industrial Metrology Germany GmbH  
Drachenloch 5, 78052 Villingen-Schwenningen**

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

*This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.*

Abbreviations used: see last page

**Page 1 of 6**

**This document is a translation. The definitive version is the original German annex to the accreditation certificate.**

**Annex to the Accreditation Certificate D-K-15030-01-00**

Calibration in the fields:

**Dimensional quantities**

**Length**

- **Roughness**
- **Form Error**
- **Contours**
- **Stylus instruments <sup>a)</sup>**
- **Length measuring instruments <sup>a)</sup>**

<sup>a)</sup> **also on-site calibrations**

Valid from: 17.03.2025

Date of issue: 17.03.2025

**Page 2 of 6**

**This document is a translation. The definitive version is the original German annex to the accreditation certificate.**

Annex to the Accreditation Certificate D-K-15030-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item  | Range  | Measurement conditions / procedure   | Expanded uncertainty of measurement   | Remarks  |
|--|--|--|---|--|
| <b>Length</b><br>Groove depth $P_t$ and $d$<br>on depth setting standards                                    | 0.15 $\mu\text{m}$ to 12 $\mu\text{m}$   | DIN EN ISO 4287:2010<br>DIN EN ISO 5436-1:2000   | $0.012 \mu\text{m} + 0.8 \cdot 10^{-3} \cdot P_t$<br>$0.012 \mu\text{m} + 0.8 \cdot 10^{-3} \cdot d$  | Groove depth $P_t$ and $d$<br>in mm  |
|  | > 12 $\mu\text{m}$ to 5500 $\mu\text{m}$   | DIN EN ISO 3274:1998<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022   | $0.022 \mu\text{m} + 0.036 \cdot 10^{-3} \cdot P_t$<br>$0.022 \mu\text{m} + 0.036 \cdot 10^{-3} \cdot d$  |  |
| Roughness on geometric standards<br>$R_z$<br>$R_{max}$ , $R_z I_{max}$<br><br>$R_{zx}(l)$<br>$R_{Sm}$        | 0.1 $\mu\text{m}$ to 3.5 $\mu\text{m}$<br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$<br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$<br><br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$<br>40 $\mu\text{m}$ to 400 $\mu\text{m}$ | DIN 4768:1990<br>DIN EN ISO 3274:1998<br>DIN EN ISO 4287:2010<br>DIN EN ISO 4288:1998<br>DIN EN ISO 16610-21: 2013<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022 | $1.5 \% \cdot R_a$<br>$1.5 \% \cdot R_z$<br>$2.0 \% \cdot R_{max}$<br>$2.0 \% \cdot R_z I_{max}$<br>$2.0 \% \cdot R_{zx}(l)$<br>1.5 $\mu\text{m}$ | If necessary, the cutoff length $\lambda_c$ can be select-ed one step shorter or up to two steps longer than specified in the standard, but not more than $\lambda_c = 2,5$ mm |
| $R_a$<br>$R_z$<br>$RP_c$<br>$Rpc$  | 0,1 $\mu\text{m}$ to 3,5 $\mu\text{m}$<br>0,5 $\mu\text{m}$ to 20 $\mu\text{m}$<br>$25 \leq RP_c \leq 150$<br>$25 \leq Rpc \leq 150$   | Steel test specification 1940<br>SEP 1940:2002<br>DIN EN 10049:2014<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022  | $5.0 \% \cdot R_a$<br>$5.0 \% \cdot R_z$<br>$2.0 \text{ cm}^{-1}$<br>$2.0 \text{ cm}^{-1}$  | Depending on the profile height other intersection line distances can be chosen (as specified)   |
| Roughness on aperiodic roughness standards<br>$R_a$<br>$R_z$<br>$R_{max}$ , $R_z I_{max}$<br><br>$R_{zx}(l)$ | 0.1 $\mu\text{m}$ to 3,5 $\mu\text{m}$<br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$<br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$<br><br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$  | DIN 4768:1990<br>DIN EN ISO 3274:1998<br>DIN EN ISO 4287:2010<br>DIN EN ISO 4288:1998<br>DIN EN ISO 16610-21: 2013<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022 | $2.5 \% \cdot R_a$<br>$3.0 \% \cdot R_z$<br>$3.5 \% \cdot R_{max}$<br>$3.5 \% \cdot R_z I_{max}$<br>$3.5 \% \cdot R_{zx}(l)$                      |  |
| $R_{pk}$<br>$R_k$<br>$R_{vk}$  | On surfaces in the range   | DIN 4776:1990<br>DIN EN ISO 13565-1:1998<br>DIN EN ISO 13565-2:1998  | $9.0 \% \cdot R_{pk}$<br>$5.0 \% \cdot R_k$<br>$8.0 \% \cdot R_{vk}$  |  |
| $Mr1$ , $R_{mr1}$<br>$Mr2$ , $R_{mr2}$   | $0.1 \mu\text{m} \leq R_a \leq 3.5 \mu\text{m}$<br>$0.5 \mu\text{m} \leq R_z \leq 20 \mu\text{m}$  | DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022<br>DIN EN ISO 16610-31:2017   | 2.0 %<br>2.0 %  | Relative measuring uncertainty relative to 100 % material ratio  |
| $R_a$<br>$R_z$<br>$RP_c$<br>$Rpc$  | 0.1 $\mu\text{m}$ to 3.5 $\mu\text{m}$<br>0.5 $\mu\text{m}$ to 20 $\mu\text{m}$<br>$25 \leq RP_c \leq 100$<br>$25 \leq Rpc \leq 100$   | Steel test specification 1940<br>SEP 1940:2002<br>DIN EN 10049:2014<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022  | $8 \% \cdot R_a$<br>$8 \% \cdot R_z$<br>$2.0 \text{ cm}^{-1}$<br>$2.0 \text{ cm}^{-1}$  | Depending on the profile height other intersection line distances can be chosen (as specified)   |

Valid from: 17.03.2025

Date of issue: 17.03.2025

Page 3 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Accreditation Certificate D-K-15030-01-00**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item   | Range  | Measurement conditions / procedure   | Expanded uncertainty of measurement  | Remarks   |
|---|--|--|--|---|
| Roughness on extra fine aperiodic roughness standards<br><i>Ra</i><br><i>Rz</i><br><i>Rmax, Rz1max</i><br><i>Rzx(l)</i> | 0.015 µm to 0.1 µm<br>0.1 µm to 0.8 µm<br>0.1 µm to 0.8 µm<br>0.1 µm to 0.8 µm | DIN 4768:1990<br>DIN EN ISO 3274:1998<br>DIN EN ISO 4287:2010<br>DIN EN ISO 4288:1998<br>DIN EN ISO 16610-21: 2013<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022 | 6 % · <i>Ra</i><br>7 % · <i>Rz</i><br>9 % · <i>Rmax</i><br>9 % · <i>Rzx(l)</i> |   |
| <i>Rpk</i><br><i>Rk</i><br><i>Rvk</i>   | On surfaces in the range   | DIN 4776:1990<br>DIN EN ISO 13565-1:1998<br>DIN EN ISO 13565-2:1998  | 10 % · <i>Rpk</i><br>6 % · <i>Rk</i><br>9 % · <i>Rvk</i>                       |   |
| <i>Mr1, Rmrk1</i><br><i>Mr2, Rmrk2</i>  | 0.015 µm ≤ <i>Ra</i> ≤ 0.1 µm<br>0.1 µm ≤ <i>Rz</i> ≤ 0.8 µm                   | DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022<br>DIN EN ISO 16610-31:2017   | 2.0 %<br>2.0 %   | Relative measuring uncertainty relative to 100 % material ratio |
| Roundness standards<br>Roundness deviation  | to 10 µm<br>> 10 µm to 20 µm   | DIN EN ISO 1101:2017<br>DIN EN ISO 12181-1:2011<br>DIN EN ISO 12181-1:2011<br>DKD-R 4-4:2018   | 0.025 µm<br>0.1 µm   | Diameter:<br>5 mm to 300 mm                                     |
| Magnification standards (flick standards)<br>Roundness deviation  | 2 µm to 20 µm<br>> 20 µm to 60 µm<br>> 60 µm to 500 µm                         |  | 0.2 µm<br>0.3 µm<br>0.5 % of measured value                                    |   |
| Cylindrical form standards<br>Roundness deviation   | to 20 µm   | DIN EN ISO 1101:2017<br>DIN EN ISO 12181-1:2011<br>DIN EN ISO 12181-1:2011<br>DKD-R 4-4:2018   | 0.1 µm   | Diameter:<br>3 mm to 300 mm<br>Length:<br>5 mm to 300 mm        |
| Straightness deviation of the generatrices<br>Length:<br>2 mm to 300 mm   | to 10 µm   |  | 0.2 µm   |   |
| Length:<br>2 mm to 100 mm   |  |  | 0.2 µm   |   |
| Length:<br>> 100 mm to 300 mm   | > 10 µm to 20 µm   |  | 0.3 µm   |   |
| Parallelism deviation of the generatrices<br>Length:<br>2 mm to 300 mm  | to 10 µm   |  | 0.3 µm   |   |
| Length:<br>2 mm to 100 mm<br>Length:<br>> 100 mm to 300 mm  | > 10 µm to 20 µm   |  | 0.3 µm<br>0.4 µm   |   |

Valid from: 17.03.2025

Date of issue: 17.03.2025

Annex to the Accreditation Certificate D-K-15030-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item   | Range  | Measurement conditions / procedure   | Expanded uncertainty of measurement   | Remarks  |
|---|--|--|---|--|
| Contour standards   |  | Substitution measurement with reference contour standard according to VDI 2629 part 1:2008<br>Procedure according to DIN EN ISO 15530-3:2012 |   |  |
| X Length<br>Lateral distances   | 5 mm to 100 mm   |  | 0.6 $\mu\text{m}$   |  |
| Z Length<br>Vertical distances  | to 10 mm   |  | 0.75 $\mu\text{m}$  |  |
| Radii   | 2 mm to 12 mm  |  | 0.75 $\mu\text{m}$  |  |
| Angles  | 40° to 135°  |  | 0.01°   |  |
| Stylus instruments according to DIN 4772:1979<br>DIN EN ISO 3274:1998<br>Groove depth $Pt$ and $d$<br>$Ra$<br>$Rz$<br>$R_{max}$ , $RzI_{max}$<br>$Rz(l)$<br>$RSm$ | 0.15 $\mu\text{m}$ to 5500 $\mu\text{m}$<br>0.015 $\mu\text{m}$ to 3.5 $\mu\text{m}$<br>0.1 $\mu\text{m}$ to 20 $\mu\text{m}$<br>0.1 $\mu\text{m}$ to 20 $\mu\text{m}$<br>0.1 $\mu\text{m}$ to 20 $\mu\text{m}$<br>40 $\mu\text{m}$ to 400 $\mu\text{m}$ | DKD-R 4-2 Blatt 2:2018<br>DIN EN ISO 12179:2020 E  | $U_{\text{normal}} + 0.01 \mu\text{m}$<br>$U_{\text{normal}} + 1 \% \cdot Ra$<br>$U_{\text{normal}} + 1 \% \cdot Rz$<br>$U_{\text{normal}} + 1 \% \cdot R_{max}$<br>$U_{\text{normal}} + 1 \% \cdot Rz(l)$<br>$U_{\text{normal}} + 1 \mu\text{m}$ | $U_{\text{normal}}$ is the measurement uncertainty of the standards used. Smaller measuring ranges for which standards are available can also be calibrated. |
| $Rpk$<br>$Rk$<br>$Rvk$  | On surfaces in the range   | DIN 4776: 1990<br>DIN EN ISO 13565-1: 1998<br>DIN EN ISO 13565-2: 1998   | $U_{\text{normal}} + 1 \% \cdot Rpk$<br>$U_{\text{normal}} + 1 \% \cdot Rk$<br>$U_{\text{normal}} + 1 \% \cdot Rvk$   |  |
| $Mr1$ , $Rmr1$<br>$Mr2$ , $Rmr2$  | $0.015 \mu\text{m} \leq Ra \leq 3.5 \mu\text{m}$<br>$0.1 \mu\text{m} \leq Rz \leq 20 \mu\text{m}$  | DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022<br>DIN EN ISO 16610-31:2017   | $U_{\text{normal}} + 1 \%$<br>$U_{\text{normal}} + 1 \%$  | Relative measuring uncertainty relative to 100 % material ratio  |
| Optoelectronic length and diameter measuring devices (Shaft measuring systems)  |  | Shadow image method<br>QMV: Kalibrierung von Wellenmessgeräten: 2021-10 (english: Calibration of shaft measuring systems 2024-12)            |   | Smaller measuring ranges for which standards are available can also be calibrated  |
| Diameter  | to 320 mm  |  | $0.4 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot d$   | $d$ = measured diameter  |
| Length  | to 1200 mm   |  | $0.5 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$   | $l$ = measured length  |

Valid from: 17.03.2025

Date of issue: 17.03.2025

**Annex to the Accreditation Certificate D-K-15030-01-00**

**On-site Calibration**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item   | Range  | Measurement conditions / procedure  | Expanded uncertainty of measurement  | Remarks   |
|---|--|---|--|---|
| <b>Length</b><br>Stylus instruments according to DIN 4772:1979<br>DIN EN ISO 3274:1998<br>Groove depth $P_t$ and $d$<br>$R_a$<br>$R_z$<br>$R_{max}$ , $R_z I_{max}$<br>$R_z x(l)$<br>$R_{Sm}$ | 0.15 $\mu\text{m}$ to 5500 $\mu\text{m}$<br>0.015 $\mu\text{m}$ to 3.5 $\mu\text{m}$<br>0.1 $\mu\text{m}$ to 20 $\mu\text{m}$<br>0.1 $\mu\text{m}$ to 20 $\mu\text{m}$<br>0.1 $\mu\text{m}$ to 20 $\mu\text{m}$<br>40 $\mu\text{m}$ to 400 $\mu\text{m}$ | DKD-R 4-2 part 2:2010<br>DIN EN ISO 12179:2020 E<br>DIN 4768: 1990<br>DIN EN ISO 3274: 1998<br>DIN EN ISO 4287: 2010<br>DIN EN ISO 4288: 1998<br>DIN EN ISO 16610-21:2013<br>DIN EN ISO 21920-2: 2022<br>DIN EN ISO 21920-3: 2022 | $U_{\text{normal}} + 0.01 \mu\text{m}$<br>$U_{\text{normal}} + 1\% \cdot R_a$<br>$U_{\text{normal}} + 1\% \cdot R_z$<br>$U_{\text{normal}} + 1\% \cdot R_{max}$<br>$U_{\text{normal}} + 1\% \cdot R_z x(l)$<br>$U_{\text{normal}} + 1 \mu\text{m}$ | $U_{\text{normal}}$ is the measurement uncertainty of the standards used.<br>Smaller measuring ranges for which standards are available can also be calibrated. |
| $R_{pk}$<br>$R_k$<br>$R_{vk}$   | On surfaces in the range<br>$0.015 \mu\text{m} \leq R_a \leq 3,5 \mu\text{m}$<br>$0.1 \mu\text{m} \leq R_z \leq 20 \mu\text{m}$  |   | $U_{\text{normal}} + 1\% \cdot R_{pk}$<br>$U_{\text{normal}} + 1\% \cdot R_k$<br>$U_{\text{normal}} + 1\% \cdot R_{vk}$  | Relative measuring uncertainty relative to 100 % material ratio   |
| $M_{r1}$ , $R_{mr1}$<br>$M_{r2}$ , $R_{mr2}$  |  |   | $U_{\text{normal}} + 1\%$<br>$U_{\text{normal}} + 1\%$   |   |
| Optoelectronic length and diameter measuring devices (Shaft measuring systems)  |  | Shadow image method<br>QMV: Kalibrierung von Wellenmessgeräten: 2021-10<br>(english: Calibration of shaft measuring systems 2024-12)  |  | Smaller measuring ranges for which standards are available can also be calibrated.  |
| Diameter  | to 320 mm  |   | $0.4 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot d$  | $d$ = measured diameter   |
| Length  | to 1200 mm   |   | $0.5 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$  | $l$ = measured length   |

**Abbreviations used:**

|       |  |
|-------|--|
| CMC   | Calibration and measurement capabilities   |
| DIN   | German Institute for Standardization e.V.  |
| DKD-R | Guidelines of the German Calibration Service (DKD), published by the Physikalisch-Technische Bundesanstalt |
| EN    | Europäische Norm   |
| ISO   | Internationale Organisation für Normung  |
| QMA   | Internal documentation of JENOPTIC Industrial Metrology Germany GmbH                                       |
| VDE   | Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.  |
| VDI   | Verein Deutscher Ingenieure e.V.   |

Valid from: 17.03.2025

Date of issue: 17.03.2025